

Gleitfadpräparation leicht gemacht

Dr. Henning Bahnemann

Endodontie Journal 1/2010

Literatur:

1 Schilder H

Cleaning and shaping the root canal.

Dent Clin North Am 1974; 18: 269–296.

2 Parashos P, Messer HH.

The diffusion of innovation in dentistry: a review using rotary nickel-titanium technology as an example.

Oral Surg Oral Med Oral Pathol Oral Radiol Endod 101, p 395-401, 2006

3 Berutti E, Negro AR, Lendini M, Pasqualini D

Influence of manual preflaring and torque on the failure rate of ProTaper rotary instruments.

J Endod 30, p 228–230, 2004

4 Patiño PV, Biedma BM, Liébana CR, Cantatore G, Bahillo JG

The influence of a manual glide path on the separation rate of NiTi rotary instruments.

J Endod 31, p 114–116, 2005

5 Gambarine, G.

Torsional and cyclic fatigue testing of ProFile NiTi rotary instruments.

Journal of Evolutionary Dentistry. v. 2, p. 4-14, 1999.

6 Blum, J.Y.; Machtou, P.; Micallef, J.P.

Location of contact areas on rotary ProFile instruments in relationship to the forces developed during mechanical preparation on extracted teeth.

Int Endod J. v. 32, p. 108-114, 1999.

7 Elio Berutti, Giuseppe Cantatore, Arnaldo Castellucci, Giorgio Chiandussi, Francesco Pera, Giuseppe Migliaretti, Damiano Pasqualini

Use of Nickel-Titanium Rotary PathFile to Create the Glide Path: Comparison With Manual Preflaring in Simulated Root Canals

J Endod, v 35, p 408-412, 2009